

#### Recombinant Human Leucine Rich Alpha-2-Glycoprotein 1 (LRG1)

Catalog No.: TP02842 50µg

#### Sequence Information

Species: Human Swiss Prot:P02750

#### Gene ID:116844

Synonyms:Leucine-rich alpha-2-glycoprotein,

HMFT1766, LRG

Residues:Thr37~Leu340

TLSPKDCQVFRSDHGSSISCQPPAEIPGYLPADTVHLAVEFFNLTHLPANLLQG

ASKLQELHLSSNGLESLSPEFLRPVPQLRVLDLTRNALTGLPPGLFQASATLDT

LVLKENQLEVLEVSWLHGLKALGHLDLSGNRLRKLPPGLLANFTLLRTLDLGEN

QLETLPPDLLRGPLQLERLHLEGNKLQVLGKDLLLPQPDLRYLFLNGNKLARVA

AGAFQGLRQLDMLDLSNNSLASVPEGLWASLGQPNWDMRDGFDISGNPWICDQN

LSDLYRWLQAQKDKMFSQNDTRCAGPEAVKGQTL

#### **Product Information**

**Source:** Recombinant expression.

Host: E.coli

Tags: N-terminal His-Tag

Subcellular Location: Secreted

**Purity:** >90%

Traits: Freeze-dried powder

**Buffer formulation:** PBS, pH7.4, containing 0.01% SKL, 1mM DTT, 5% Trehalose and Proclin300.

Original Concentration: 200µg/mL

Applications: Positive Control; Immunogen; SDS-PAGE; WB.

(May be suitable for use in other assays to be determined by the end user.)

Predicted isoelectric point: 6.3

Predicted Molecular Mass: 37.4kDa

Accurate Molecular Mass: 37kDa as determined by SDS-PAGE reducing conditions.

# [ <u>USAGE</u> ]

Reconstitute in  $ddH_2O$  to a concentration of 0.1-0.5 mg/mL. Do not vortex.

# [ STORAGE AND STABILITY ]

Storage: Avoid repeated freeze/thaw cycles.



Store at 2-8°C for one month.

Aliquot and store at -80°C for 12 months.

**Stability Test:** The thermal stability is described by the loss rate. The loss rate was determined by accelerated thermal degradation test, that is, incubate the protein at 37°C for 48h, and no obvious degradation and precipitation were observed. The loss rate is less than 5% within the expiration date under appropriate storage condition.

### [IDENTIFICATION]




Figure 1. Gene Sequencing (Extract)

KDa	
180 130	
100	
70	-
55	
40	
35	
35 25	
15	- market and the second s
10	the second s

Figure 2. SDS-PAGE

# [ IMPORTANT NOTE ]

The kit is designed for in vitro and research use only, we will not be responsible for any issue if the kit was used in clinical diagnostic or any other procedures.